



**URGENT MEDICAL DEVICE
RECALL**

**T3030 - LEGENCY REMEDIES
SODIUM CHLORIDE 0.9%
IRRIGATION (AMPOULES) /
30ML / BOX OF 30**

Batch number: 2386055E

ARTG: 409695

YES Medical, following consultation with the Therapeutic Goods Administration, is recalling all batches of T3030 - LEGENCY REMEDIES SODIUM CHLORIDE 0.9% IRRIGATION (AMPOULES) / 30ML / BOX OF 30 due to suspected contamination with *Ralstonia pickettii*.

If you have any stock of T3030 - LEGENCY REMEDIES SODIUM CHLORIDE 0.9% IRRIGATION (AMPOULES) / 30ML / BOX OF 30, do not use it. Return it to the place of purchase for a refund.

Ralstonia picketti is a bacteria commonly found in the natural environment. It rarely causes infection in humans but may cause infection in people with weakened immune systems. Anyone who is concerned in any way about the use of this product should consult their doctor.

YES Medical sincerely regrets any inconvenience to their customers.

Customer Enquiries: 1800 10 10 33